


## Operating Instructions



<b>1</b>	<b>General Information</b> .....	<b>2</b>
1.1	Intended Use.....	2
1.2	General Safety Notes.....	2
1.3	Conformity with EU Regulations.....	2
1.4	ATEX Approval.....	2
<b>2</b>	<b>Transportation and Storage</b> .....	<b>2</b>
<b>3</b>	<b>Installation and Commissioning</b> .....	<b>2</b>
3.1	Mechanical Installation .....	3
3.2	Electrical Connection .....	3
3.3	Adjusting of contacts .....	4
3.4	Additional requirements relating to type BN4xxx pressure switches and type BP4xxx pressure limiters.....	4
<b>4</b>	<b>Operation</b> .....	<b>5</b>
4.1	Zero-point correction .....	5
4.2	Maintenance / Service.....	5
<b>5</b>	<b>Disassembly</b> .....	<b>5</b>

## 1 General Information

This document contains necessary information for the proper installation and use of this device. In addition to this instruction, be sure to observe all statutory requirements, applicable standards, the additional technical specifications on the accompanying data sheet (see [www.labom.com](http://www.labom.com)) as well as the specifications indicated on the type plate.

### 1.1 Intended Use

The device is intended to measure pressure of gases, vapors and liquids as specified in the data sheet.

### 1.2 General Safety Notes

The installation, set up, service or disassembly of this device must only be done by trained, qualified personnel using suitable equipment and authorized to do so.



#### Warning

Media can escape if unsuitable devices are used or if the installation is not correct.

Danger of severe injury or damage

- Ensure that the device is suitable for the process and undamaged.

### 1.3 Conformity with EU Regulations

The CE-marking on the device certifies its compliance with the applicable EU Directives for placing products on the market within the European Union.

You find the complete EU Declaration of Conformity (document no. KE\_001 (für BE4xxx, BR42xx, BG3xxx), KE\_012 (für BP4xx0), KE\_014 (für BN4xx0) or KE\_039 (für BE2xxx, BF2xxx, BG2xxx)) at [www.labom.com](http://www.labom.com).

### 1.4 ATEX Approval

If you purchased a device with ATEX approval, please refer to the accompanying document XA\_013, XA\_014 or XA\_021 for ATEX-relevant information.

## 2 Transportation and Storage

Store and transport the device only under clean and dry conditions preferably in the original packaging. Avoid exposure to shocks and excessive vibrations.

Permissible storage temperature: -40...70 °C

S3 models with case filling

Permissible storage temperature: -20...60 °C

## 3 Installation and Commissioning

Ensure that the device is suitable for the intended application with respect to pressure range, overpressure limit, media compatibility, temperature range and process connection.


### 3.1 Mechanical Installation


Use gaskets, if required, that are suitable for the process connection and resistant to the media.

Before starting operation, check the process connection carefully for leaks under pressure.

Install safety pressure gauges in the line with DIN EN 837-1 S1 and S3 so that the exhaust equipment can freely discharge to the rear in the event of a fault or malfunction. Provide for corresponding openings if mounting the unit on the wall.

Use the spanner flats to screw in the device.

Measuring devices where oil or grease residues in the measuring element are not permitted are marked on the scale with the oil can  icon.

Measuring devices for use with oxygen are marked on the scale with the oil can  icon and the word "oxygen" directly below this.

Seal cylindrical screw threads by fitting a sealing washer to the front face of the gasket. Seal taper screw threads by screwing tight; typically a gasket material is applied to the male thread. Make sure that the set of gasket you use are correct and in perfect condition when mounting the instrument; non-matching gasket can cause faults.

Vent the pressure gauge after installation using the vent valve (Figure 1).

The devices are supplied with a closed vent valve (CLOSED position). Moving the red lever to the OPEN position equalises the pressure level inside and outside the case (Figure 2).



Figure 1: Vent valve




Figure 2: Move valve

### 3.2 Electrical Connection

Complete the mechanical installation before you connect the device electrically.

Set up all electrical connections while the voltage supply is switched off.

### 3.3 Adjusting of contacts

Setting the contact can be done by using the adjustment lock in the centre of the sight glass. Just press down the separate or permanently mounted key (Figure 3) until the pointer of the lock catches the adjusting pin of the contact (Figure 4).



Figure 3: Key




Figure 4: Setting the contact

The set pointer can be adjusted over the entire scale range. It must be ensured that the contact will be set in a clockwise direction only. If the contact is accidentally set behind the desired value, turn back the contact at least 5% below the desired value and make the setting again in a clockwise direction.

The functionality and thus the switching reliability of the built-in switch contacts is guaranteed within the entire nominal range. Following valid standards though the accuracy stated in the data sheet is assured for switching points that are set between 10% and 90% of the nominal range only.

Operating the contact outside of the scale range is not permitted.

You can find additional information on contacts in the document TA\_039, available on our website at [www.labom.com](http://www.labom.com).

### 3.4 Additional requirements relating to type BN4xxx pressure switches and type BP4xxx pressure limiters

The evaluation unit and downstream control current circuit are not part of our scope of supply. The following evaluation units conform to the requirements of the Association of Technical Inspection Agencies (VdTÜV) Fact Sheet No. 100.

Pepperl+Fuchs, switching amplifier:

- Type KHA6-SH-Ex1, PTB 00 ATEX 2043
- Type KFD2-SH-Ex1, PTB 00 ATEX 2042

The use of alternative evaluation units is within the responsibility of the operator. The operator is also responsible for the control current circuit and the locking switch required with pressure limiters, that comply with the Association of Technical Inspection Agencies (VdTÜV) Fact Sheet No. 100.

## 4 Operation

During operation, take care that the device remains within its intended pressure and temperature ranges. No other monitoring is necessary.

The permissible ambient and media temperature depends on the type of device and its design. This information can be found in the relevant data sheet.

Restricted ambient conditions may apply in explosion protected environments (see XA\_013, XA\_014 or XA\_021).

### 4.1 Zero-point correction

Small measuring errors or deviations caused by difference in level between pressure gauge and point of measurement can be corrected on measuring devices with micro control position pointers. To do so turn the adjusting screw on the pointer hub (see figure 5).

Similarly, a displacement of the zero point caused by use and long service life can be corrected if necessary.



Figure 5: Zero-point correction

### 4.2 Maintenance / Service

When properly installed in accordance with applicable specifications, this device is maintenance-free. However, we recommend an annual recalibration of the device.

## 5 Disassembly

When measuring hot media, make sure that the device has cooled down prior to any dismounting or wear appropriate protective clothing to avoid burns.

Switch off the power supply to the device before disconnecting the electrical connections. Once this is done, the device may be mechanically removed.



### Warning

Opening pressurized lines might cause severe injuries.

Danger of severe injuries or damage

- Relieve the process pressure before attempting to remove the device. Shut off the pressure supply for all feed lines to the device and relieve the pressure in them.



### Warning

Hazardous deposits and residues might remain on opened process connections and removed devices.

Danger of injury

- After the device has been removed, seal off the measuring point and mark the open process connection accordingly. Consider a possible danger due to residues when handling the removed device.